



These products are sold for Research Use Only; not for use in diagnostic procedures

# Workflow Summary

- 01 Using gloved hands, remove the Master Mixes from the freezer. Allow the tubes to thaw; then gently vortex to mix.
- 02 In a containment hood or dead air box, pipette 45 µL of Master Mix into individual wells of a PCR plate (use a different indexed Master Mix for each sample and control).
- 03 Add 0.2 µL EagleTaq™ DNA polymerase to each Master Mix.
- 04 Add 5 µL of DNA (at a minimum of 10 ng/µL) from the unknown samples and controls to wells containing the respective Master Mix reactions then pipette up and down 5-10 times to mix.
- 05 Add 5 µL of molecular biology grade water to the well containing the Master Mix for the no template control then pipette up and down 5-10 times to mix.
- 06 Seal the plate and amplify target DNA using the following thermal cycler program:  

Standardized Program			
Step	Temperature	Time	Cycle
1	95 °C	7 minutes	1
2	95 °C	45 seconds	29x
3	60 °C	45 seconds	
4	72 °C	90 seconds	
5	72 °C	10 minutes	1
6	15 °C	∞	1
- 07 Purify the PCR products using the Agencourt® AMPure® XP PCR Purification system. Add 90 µL of particles to each 50 µL reaction; elute DNA in 40 µL of eluant.
- 08 Quantify amplicons with an appropriate method (e.g. Bioanalyzer® 2100 or LabChip® GX).
- 09 Create the library by combining 4 nM of each amplicon in a single tube (do not include the no template control).
- 10 Dilute the library.
- 11 Prepare templates using the OT2 or Ion Chef system.  
\*The OT2 system is used to perform an emulsion PCR to create template-positive ion sphere particles (ISPs) that must be further enriched.

	Ion PGM	Ion S5
OT2	Ion PGM Hi-Q OT2 Kit or Ion PGM Hi-Q view OT2 Kit or OT2 200 Kit if only testing TRG	Ion 520 & Ion 530 Kit – OT2
Ion Chef	Not applicable	Ion 510 & Ion 520 & Ion 530 Kit – Chef

- 12 \*OT2 templates: Enrich template-positive ISPs with the Ion OneTouch(tm) ES.
- 13 Initialize the S5 or PGM. Load S5 [Ion 520, Ion 530] or PGM chip [316 v2 BC, 318 v2 BC] with prepared templates.
- 14 Create a Planned Run using the Torrent Browser.
- 15 Start the S5/PGM run.
- 16 Analyze and visualize the acquired data using the LymphoTrack Software for the S5/PGM.

## Ordering Information

CATALOG #	PRODUCTS	QUANTITY
7-121-0057	LymphoTrack® IGH FR1/2/3 Assay - S5/PGM™	12 indices - 5 sequencing reactions each
7-121-0007	LymphoTrack® IGH FR1 Assay - S5/PGM™	12 indices - 5 sequencing reactions each
7-121-0037	LymphoTrack® IGH FR2 Assay - S5/PGM™	12 indices - 5 sequencing reactions each
7-121-0047	LymphoTrack® IGH FR3 Assay - S5/PGM™	12 indices - 5 sequencing reactions each
7-122-0007	LymphoTrack® IGK Assay - S5/PGM™	12 indices - 5 sequencing reactions each
7-227-0007	LymphoTrack® TRG Assay - S5/PGM™	12 indices - 5 sequencing reactions each
7-500-0007	LymphoTrack® Software - S5/PGM™	1 CD complimentary with purchase
7-500-0008	LymphoTrack® MRD Software	1 CD complimentary with purchase

**Storage Conditions:** -85 °C to -65 °C (DNA controls may be separated from kits and stored at 2 °C to 8 °C).

These products are sold for Research Use Only; not for use in diagnostic procedures.

©2021 Invivoscribe, Inc. (Invivoscribe®). All rights reserved. The trademarks mentioned herein are the property of Invivoscribe and/or its affiliates, or (as to the trademarks of others used herein) their respective owners. Hi-Q™, ONETOUGH™, and Ion Chef, Ion 316™, Ion 318™, Ion 510™, Ion 520™ and Ion 530™ are trademarks of Thermo Fisher Scientific or its subsidiaries. AGILENT® and BIOANALYZER® are registered trademarks of Agilent Technologies, Inc. AMPURE® is a registered trademark of Beckman Coulter, Inc. LABCHIP® is a registered trademark of Perkin Elmer. EAGLETAQ™ is a trademark of Roche.

M-0043 Rev 02 May 2021