

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Tracer iodination date: 22 January 2026

Kit lot number: 3INK366

Kit expiry date: 20 March 2026

125I - Insulin	Lot No.	IT366R	Expiry Date:	20 MAR 2026
Assay Buffer	Lot No.	MIAB172	Expiry Date:	07 JUN 2027
Anti-human IgG	Lot No.	201PG	Expiry Date:	21 APR 2027
Calibrators	Lot No.	IMS40E	Expiry Date:	26 FEB 2027
Negative Control	Lot No.	IN133	Expiry Date:	16 JUL 2026
Kit Controls	Lot No.	IMCa40E/IMCb40E	Expiry Date:	26 FEB 2027

Total cpm in 25 µl 125I - Insulin 27376

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.7		
INS-Ab calibrator:			
0.4 u/mL	2.3		
1 u/mL	4.4		
10 u/mL	19.7		
50 u/mL	59.4		
Control I	4.4	1.00	0.45 - 1.45
Control II	23.3	12.2	6 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 28 Jan 2026

Authorised by: DR G Flood

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Senior Scientist

Date: 28th JAN 2026