

Title: **Quality Control Record Sheet**
MuSK Ab Assay QC Data

Prepared by: G FLOOD
Approved by: J BRADBURY

Document: wi/qcrs/0.30a
Revision: 0
Issue date: 20.10.2021
Page 1 of 1

Kit Lot Number: KMK280
¹²⁵I-labelled MuSK: MK280
Reconstitution Buffer: 179B
Anti Human IgG: 185Gab
Precipitation Enhancer: PE204
Wash Solution: 477W
Negative Control: MKN45c
Positive Control: MKP45g

Kit Expiry Date: 21 JULY 2026
Expiry Date: 21 JULY 2026
Expiry Date: 29 NOVEMBER 2026
Expiry Date: 28 FEBRUARY 2027
Expiry Date: 28 JANUARY 2027
Expiry Date: 17 DECEMBER 2026
Expiry Date: 30 NOVEMBER 2026
Expiry Date: 30 NOVEMBER 2026

Specific activity of ¹²⁵I-labelled MuSK: 2176 Ci/mmol (39.6 μ Ci/ μ g)

Total cpm in 50 μ L of labelled MuSK: 53936

MuSK labelling date: 26 MAY 2026

% Counter Efficiency: 75.8

¹²⁵I Decay Table:

QC date (weeks after labelling date)	Decay factor
up to 1 week	1.0
+1 - 2 weeks	1.1
+2 - 3 weeks	1.2
+3 - 4 weeks	1.3
+4 - 5 weeks	1.4
+ 5 - 6 weeks	1.5
+ 6 - 7 weeks	1.6
+ 7 - 8 weeks	1.75

Positive Control	Mean cpm	Nmol/L	Range (nmol/L)
Negative Control	363	-	-
Positive Control	16024	0.8552	0.82 $\pm$ 0.41

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 29 MAY 2026
Performed by: Jonathon Davies
Signature: [Signature]
Position: Senior Scientist

Authorised by: DR G FLOOD
Signature: [Signature]
Position: Head of Quality Control
Date: 29th MAY 2026