

Title: **Quality Control Record Sheet**
21-OH Ab ELISA Kit QC Data

Document: **wi/qcrs/40**
Revision: **2**
Date: **10.10.18**
Page **1 of 1**

Prepared by: G. Flood
Approved by: W. Dowler

Kit lot number: K21E30C Expiry date: 03 July 2023

Reagent Lot Numbers:

Coated Wells: <u>21E30</u>	Expiry date: <u>03 July 2023</u>
21-OH-Biotin: <u>21EB22c</u>	Expiry date: <u>03 July 2023</u>
Recon. Buffer for 21-OH-Biotin: <u>21ER25</u>	Expiry date: <u>19 July 2023</u>
SA-POD: <u>GEP499</u>	Expiry date: <u>28 December 2023</u>
Diluent for SA-POD: <u>2RPD78</u>	Expiry date: <u>21 April 2024</u>
Reaction Enhancer: <u>IES115</u>	Expiry date: <u>10 February 2024</u>
Peroxidase Substrate: <u>PTS126</u>	Expiry date: <u>23 November 2023</u>
Stop Solution: <u>HESP117</u>	Expiry date: <u>10 March 2024</u>
Concentrated Wash: <u>CW307</u>	Expiry date: <u>11 July 2024</u>
Calibrators: <u>21C30c</u>	Expiry date: <u>03 July 2023</u>
Reference Preparation: <u>21ERP30c</u>	Expiry date: <u>03 July 2023</u>
Negative Control: <u>21NC30c</u>	Expiry date: <u>03 July 2023</u>
Positive Controls: <u>21PCa/b30c</u>	Expiry date: <u>03 July 2023</u>

Curve and Controls: Incubation temp: 21°C

	Absorbance at 450 nm	Absorbance at 405 nm
0.3 units per mL	0.189	0.055
1 units per mL	0.620	0.182
10 units per mL	1.943	0.574
100 units per mL	4.021	1.183
	(derived from Abs at 405nm)	
Reference Preparation	0.548	0.159
Negative Control	0.024	0.007
Positive Control I	0.705	0.206
Positive Control II	1.850	0.546

Control	Concentration at 450 nm (units per mL)	Concentration at 405 nm (units per mL)	Range (units per mL)	Index Value at 450 nm	Index Value at 405 nm	Range (index value)
Negative	<0.3	<0.3	<0.3	<45	<45	<45
Positive I	1.21	1.20	0.8 – 1.9	128.6	129.6	70 - 170
Positive II	8.83	8.81	5 - 13	337.6	343.4	190 - 430

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay performed by: Hannah Sanyaolu
Signature: [Signature]
Position: Senior Scientist
Date: 12 July 2022

Authorised by: DR G Flood
Signature: [Signature]
Position: HEAD OF QC
Date: 21ST SEP 2022