Title:

Quality Control Record Sheet GAD Ab ELISA Kit QC Data

Prepared by: G Flood

Approved by: L Carr

Document: Revision:

wi/qcrs/0.27

Date:

10/10/14

Page

1 of 1

Kit lot number: KGE411A Kit expiry date: 07 February 2022 Kit manufacture date: 08 February 2021 GAD₆₅ Coated Wells: Lot No. GE446 Expiry Date 09 May 2022 GAD₆₅ -Biotin: Lot No. GEB378 Expiry Date 08 May 2022 Recon. Buffer for GAD₆₅-Biotin: Lot No. GBB157 Expiry Date 10 May 2022 Streptavidin Peroxidase: Lot No. GEP395 Expiry Date 03 May 2022 Diluent for SA-POD: Lot No. PB161 Expiry Date 18 October 2022 Peroxidase Substrate: Lot No. PTS173H Expiry Date 19 November 2022 Stop Solution: Lot No. HESP107 Expiry Date 15 September 2022 Concentrated Wash Solution: Lot No. CW262 Expiry Date 19 April 2023 Calibrators: Lot No. GEC413a Expiry Date 07 February 2022 Controls: Lot No. <u>GENC413a/GEPC413a</u> Expiry Date <u>07 February 2022</u>

Calibrator Curve and Controls: Incubation temp: °C **Calibrator Concentration** Absorbance at Absorbance at (units per mL) 450 nm 405 nm Negative control 0.016 0.004 5 0.206 0.061 18 0.603 0.177 35 1.155 0.342 120 2.724 0.805 250 3.752 1.139 2000 >4 1.850

Control	Absorbance at 450 nm	Absorbance at 405 nm	Concentration at 450 nm (units per mL)	Concentration at 405 nm (units per mL)	Range (units per mL)
Positive	1.461	0.432	46.2	46.5	F1 : 10
			10.2	40.5	51 ± 10

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay performed by:	Xing		
Signature:	2 houghing leng		
Position:	Laboratory Technician		
Date:	10 December 2020		
Authorised by:	DR & GDOD		
Signature:	5 Marsh		
Position:	MEAD OF OC		
Date:	2187 JAN 2021		