

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Tracer iodination date: 18 April 2024

Kit lot number: 3INK343

Kit expiry date: 14 June 2024

125I - Insulin	Lot No.	IT343R	Expiry Date:	14 JUN 2024
Assay Buffer	Lot No.	MIAB152	Expiry Date:	25 AUG 2025
Anti-human IgG	Lot No.	188PG	Expiry Date:	14 JUL 2025
Calibrators	Lot No.	IMS38C	Expiry Date:	26 FEB 2025
Negative Control	Lot No.	IN132	Expiry Date:	04 JUL 2025
Kit Controls	Lot No.	IMCa38C/IMCb38C	Expiry Date:	26 FEB 2025

Total cpm in 25 μ l 125I - Insulin 29942

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.7		
INS-Ab calibrator:			
0.4 u/mL	2.5		
1 u/mL	4.7		
10 u/mL	19.0		
50 u/mL	62.8		
Control I	4.9	1.1	0.5 - 1.7
Control II	22.9	12.2	8 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 24 Apr 2024

Performed by: Fiona Ryrie

Signature: Fiona Ryrie

Position: Scientist

Authorised by: DR G Flood

Signature: [Signature]

Position: Head of Quality Control

Date: 25th APR 2024