

Title: **Quality Control Record Sheet**
Insulin Ab 3 QC Data

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 16 April 2025

Kit lot number: 3INK356


Kit expiry date: 12 June 2025


125I - Insulin	Lot No. IT356R	Expiry Date: 12 JUN 2025
Assay Buffer	Lot No. MIAB163	Expiry Date: 24 AUG 2026
Anti-human IgG	Lot No. 196PG	Expiry Date: 03 AUG 2026
Calibrators	Lot No. IMS40	Expiry Date: 26 FEB 2027
Negative Control	Lot No. IN132	Expiry Date: 04 JUL 2025
Kit Controls	Lot No. IMCa40/IMCb40	Expiry Date: 26 FEB 2027

Total cpm in 25 µl 125I - Insulin 28785

<u>Sample</u>	<u>% 125I - Insulin</u> <u>bound</u>	<u>u/mL</u>	<u>Range</u> <u>u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	2.0		
1 u/mL	3.7		
10 u/mL	15.3		
50 u/mL	56.9		
Control I	3.4	0.81	0.45 – 1.45
Control II	18.1	11.7	6 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 23 Apr 2025
Performed by: J DAVIES
Signature: 
Position: Senior Scientist

Authorised by: DR G Flood
Signature: 
Position: Head of Quality Control
Date: 23 RD APR 2025