

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**
 Revision: **2**
 Date: **09.01.14**
 Page **1 of 1**

Prepared by: Geoff Flood
 Approved by: Janet Bradbury

Tracer iodination date: 08 October 2020

Kit lot number: 3INK297


Kit expiry date: 04 December 2020



125I - Insulin	Lot No. IT297R	Expiry Date: 04 DEC 2020
Assay Buffer	Lot No. MIAB122	Expiry Date: 01 FEB 2022
Anti-human IgG	Lot No. 171PG	Expiry Date: 24 FEB 2022
Calibrators	Lot No. IMS35	Expiry Date: 02 SEP 2022
Negative Control	Lot No. IN127	Expiry Date: 10 APR 2021
Kit Controls	Lot No. IMCa35/IMCb35	Expiry Date: 02 SEP 2022

Total cpm in 25 µl 125I - Insulin 28142

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>0.9</u>		
INS-Ab calibrator:			
0.4 u/mL	2.0		
1 u/mL	4.0		
10 u/mL	16.2		
50 u/mL	54.0		
Control I	3.7	0.85	0.5 – 1.5
Control II	19.9	12.4	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 14 Oct 2020
 Performed by: Becky Coles
 Signature: 
 Position: Senior Scientist

Authorised by: 
 Signature: 
 Position: Head of Quality Control
 Date: 19th OCT 2020