

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 03 November 2022

Kit lot number: ^{AW}K3INK324

Kit expiry date: 30 December 2022

125I - Insulin	Lot No. IT324R	Expiry Date: 30 DEC 2022
Assay Buffer	Lot No. MIAB138	Expiry Date: 11 FEB 2024
Anti-human IgG	Lot No. 180PG	Expiry Date: 27 DEC 2023
Calibrators	Lot No. IMS37A	Expiry Date: 25 MAY 2024
Negative Control	Lot No. IN130	Expiry Date: 10 OCT 2023
Kit Controls	Lot No. IMCa37A/IMCb37A	Expiry Date: 25 MAY 2024

Total cpm in 25 μ l 125I - Insulin 27650

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	1.0		
INS-Ab calibrator:			
0.4 u/mL	2.6		
1 u/mL	4.9		
10 u/mL	19.4		
50 u/mL	58.2		
Control I	4.3	0.78	0.5 - 1.6
Control II	21.2	11.1	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 10 Nov 2022

Performed by: A Huckle

Signature: A. Huckle

Position: Snr Technician

Authorised by: DR G Flood

Signature: [Signature]

Position: Head of Quality Control

Date: 10TH NOV 2022