

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood  
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b  
Revision: 2  
Date: 09.01.14  
Page 1 of 1

Tracer iodination date: 02 November 2023

Kit lot number: 3INK337


Kit expiry date: 29 December 2023


125I - Insulin	Lot No.	IT337R	Expiry Date:	29 DEC 2023
Assay Buffer	Lot No.	MIAB147	Expiry Date:	17 FEB 2025
Anti-human IgG	Lot No.	185PG	Expiry Date:	21 FEB 2025
Calibrators	Lot No.	IMS38B	Expiry Date:	26 FEB 2025
Negative Control	Lot No.	IN131	Expiry Date:	06 JUL 2024
Kit Controls	Lot No.	IMCa38B/IMCb38B	Expiry Date:	26 FEB 2025

Total cpm in 25 µl 125I - Insulin 28184

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	1.9		
1 u/mL	4.0		
10 u/mL	18.0		
50 u/mL	58.7		
Control I	4.0	1.00	0.5 – 1.7
Control II	21.2	11.9	8 – 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 08 Nov 2023  
 Performed by: J DAVIES  
 Signature:   
 Position: Senior Scientist

Authorised by: DR G Flood  
 Signature:   
 Position: Head of Quality Control  
 Date: 8<sup>th</sup> NOV 2023