

Title:

Quality Control Record Sheet Insulin Ab 3 QC Data

Document:

wi/qcrs/0.08b

Revision:

2

Prepared by: Geoff Flood

Date:

09.01.14

Approved by: Janet Bradbury

Page

1 of 1



Tracer iodination date: 17 June 2021Kit lot number: 3INK306Kit expiry date: 13 August 2021

125I - Insulin	Lot No. IT306R	Expiry Date: 13 AUG 2021
Assay Buffer	Lot No. MIAB127	Expiry Date: ⁰⁶ 03 SEP 2022
Anti-human IgG	Lot No. 170PG	Expiry Date: 18 DEC 2021
Calibrators	Lot No. IMS35C	Expiry Date: 02 SEP 2022
Negative Control	Lot No. IN128	Expiry Date: 26 JUN 2022
Kit Controls	Lot No. IMCa/b35C	Expiry Date: 02 SEP 2022

Total cpm in 25 µl 125I - Insulin 29383

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>1.0</u>		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	4.3		
10 u/mL	19.1		
50 u/mL	56.4		
Control I	3.9	0.84	0.5 – 1.5
Control II	24.2	13.3	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 22 Jun 2021Authorised by: DR G FloodPerformed by: A HuckleSignature: Signature: Position: Head of Quality ControlPosition: Snr TechnicianDate: 23RD JUN 2021