

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 06 January 2022

Kit lot number: 3INK313

Kit expiry date: 04 March 2022

125I - Insulin	Lot No. IT313R	Expiry Date: 04 MAR 2022
Assay Buffer	Lot No. MIAB131	Expiry Date: 12 MAR 2023
Anti-human IgG	Lot No. 176PG	Expiry Date: 20 MAR 2023
Calibrators	Lot No. IMS36B	Expiry Date: 15 JUN 2023
Negative Control	Lot No. IN128	Expiry Date: 26 JUN 2022
Kit Controls	Lot No. IMCa/b36B	Expiry Date: 15 JUN 2023

Total cpm in 25 µl 125I - Insulin 29597

Sample	% 125I - Insulin bound	u/mL	Range u/mL
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	2.1		
1 u/mL	3.4		
10 u/mL	15.8		
50 u/mL	56.9		
Control I	3.5	1.1	0.5 - 1.4
Control II	19.6	12.3	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 11 January 2022

Authorised by: DR G Flood

Performed by: J Bradbury

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Head of Operations

Date: 12th JAN 2022