

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 27 January 2022

Kit lot number: 3INK314

Kit expiry date: 25 March 2022

125I - Insulin	Lot No. IT314R	Expiry Date: 25 MAR 2022
Assay Buffer	Lot No. MIAB132	Expiry Date: 22 APR 2023
Anti-human IgG	Lot No. 177PG	Expiry Date: 30 APR 2023
Calibrators	Lot No. IMS36B	Expiry Date: 15 JUN 2023
Negative Control	Lot No. IN128	Expiry Date: 26 JUN 2022
Kit Controls	Lot No. IMCa36B/IMCb36B	Expiry Date: 15 JUN 2023

Total cpm in 25 µl 125I - Insulin 30786

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.1		
1 u/mL	3.6		
10 u/mL	15.7		
50 u/mL	52.9		
Control I	3.3	0.81	0.5 - 1.4
Control II	18.3	11.7	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 01 Feb 2022

Authorised by: DR G Flood

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Scientist

Date: 2ND FEB 2022