

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**
Revision: **2**
Date: **09.01.14**
Page **1 of 1**

Prepared by: **Geoff Flood**
Approved by: **Janet Bradbury**

Tracer iodination date: 24 February 2022

Kit lot number: 3INK315

Kit expiry date: 22 April 2022

125I - Insulin	Lot No.	IT315R	Expiry Date:	22 APR 2022
Assay Buffer	Lot No.	MIAB132	Expiry Date:	22 APR 2023
Anti-human IgG	Lot No.	177PG	Expiry Date:	30 APR 2023
Calibrators	Lot No.	IMS36B	Expiry Date:	15 JUN 2023
Negative Control	Lot No.	IN128	Expiry Date:	26 JUN 2022
Kit Controls	Lot No.	IMCa36B/IMCb36B	Expiry Date:	15 JUN 2023

Total cpm in 25 μ l ¹²⁵I - Insulin 29355

<u>Sample</u>	<u>% ¹²⁵I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	1.0		
INS-Ab calibrator:			
0.4 u/mL	2.5		
1 u/mL	4.4		
10 u/mL	20.4		
50 u/mL	61.7		
Control I	4.3	0.95	0.5 - 1.4
Control II	24.9	12.6	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 01 Mar 2022

Performed by: F Ryrie

Signature: *Fiona Ryrie*

Position: Scientist

Authorised by: *DR G FLOOD*

Signature: *G Flood*

Position: Head of Quality Control

Date: 3RD MAR 2022