

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 21 April 2022

Kit lot number: 3INK317

Kit expiry date: 17 June 2022

125I - Insulin	Lot No. IT317R	Expiry Date: 17 JUN 2022
Assay Buffer	Lot No. MIAB134	Expiry Date: 02 AUG 2023
Anti-human IgG	Lot No. 178PG	Expiry Date: 07 SEP 2023
Calibrators	Lot No. IMS36A	Expiry Date: 15 JUN 2023
Negative Control	Lot No. IN129	Expiry Date: 24 OCT 2022
Kit Controls	Lot No. IMCa36A/IMCb36A	Expiry Date: 15 JUN 2023


Total cpm in 25 µl 125I - Insulin 29065

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	1.0		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	3.6		
10 u/mL	16.9		
50 u/mL	52.1		
Control I	3.8	1.1	0.5 - 1.4
Control II	19.1	11.5	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.


Assay Date: 26 Apr 2022

Performed by: J DAVIES

Signature: 

Position: Scientist

Authorised by: DR G Flood

Signature: 

Position: Head of Quality Control

Date: 27th APR 2022