

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**
 Revision: **2**
 Date: **09.01.14**
 Page **1 of 1**

Prepared by: **Geoff Flood**
 Approved by: **Janet Bradbury**

Tracer iodination date: 19 May 2022

Kit lot number: 3INK318B

Kit expiry date: 15 July 2022

125I - Insulin	Lot No.	IT318R	Expiry Date:	15 JUL 2022
Assay Buffer	Lot No.	MIAB135	Expiry Date:	10 SEP 2023
Anti-human IgG	Lot No.	178PG	Expiry Date:	07 SEP 2023
Calibrators	Lot No.	IMS36E	Expiry Date:	15 JUN 2023
Negative Control	Lot No.	IN129	Expiry Date:	24 OCT 2022
Kit Controls	Lot No.	IMCa36E/IMCb36E	Expiry Date:	15 JUN 2023

Total cpm in 25 µl 125I - Insulin 28344

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.4		
1 u/mL	4.3		
10 u/mL	17.7		
50 u/mL	55.4		
Control I	3.6	0.71	0.5 - 1.4
Control II	20.0	11.5	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.


Assay Date: 24 May 2022

Performed by: J DAVIES

Signature: 

Position: Scientist

Authorised by: DR G Flood

Signature: 

Position: Head of Quality Control

Date: 25TH MAY 2022