

Title: **Quality Control Record Sheet**  
**Insulin Ab 3 QC Data**

Prepared by: Geoff Flood  
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b  
Revision: 2  
Date: 09.01.14  
Page 1 of 1

Tracer iodination date: 16 June 2022

Kit lot number: 3INK319

Kit expiry date: 12 August 2022

125I - Insulin	Lot No. IT319R	Expiry Date: 12 AUG 2022
Assay Buffer	Lot No. MIAB135	Expiry Date: 10 SEP 2023
Anti-human IgG	Lot No. 178PG	Expiry Date: 07 SEP 2023
Calibrators	Lot No. IMS36E	Expiry Date: 15 JUN 2023
Negative Control	Lot No. IN129	Expiry Date: 24 OCT 2022
Kit Controls	Lot No. IMCa36E/IMCb36E	Expiry Date: 15 JUN 2023

Total cpm in 25 µl 125I - Insulin 30182

<u>Sample</u>	<u>% 125I - Insulin</u> <u>bound</u>	<u>u/mL</u>	<u>Range</u> <u>u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	3.8		
10 u/mL	14.6		
50 u/mL	53.6		
Control I	3.4	0.75	0.5 - 1.4
Control II	19.1	12.9	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.


Assay Date: 21 Jun 2022

Performed by: J DAVIES

Signature: 

Position: Scientist

Authorised by: DR G Flood

Signature: 

Position: Head of Quality Control

Date: 22<sup>nd</sup> JUN 2022