

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**

Revision: **2**

Prepared by: **Geoff Flood**

Date: **09.01.14**

Approved by: **Janet Bradbury**

Page **1 of 1**

Tracer iodination date: 14 July 2022

Kit lot number: 3INK320

Kit expiry date: 09 September 2022

125I - Insulin	Lot No.	IT320R	Expiry Date:	09 SEP 2022
Assay Buffer	Lot No.	MIAB135	Expiry Date:	10 SEP 2023
Anti-human IgG	Lot No.	179PG	Expiry Date:	05 NOV 2023
Calibrators	Lot No.	IMS36E	Expiry Date:	15 JUN 2023
Negative Control	Lot No.	IN129	Expiry Date:	24 OCT 2022
Kit Controls	Lot No.	IMCa36E/IMCb36E	Expiry Date:	15 JUN 2023

Total cpm in 25 µl 125I - Insulin 28300

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.5		
1 u/mL	3.5		
10 u/mL	15.6		
50 u/mL	54.8		
Control I	3.6	1.2	0.5 - 1.4
Control II	19.1	12.2	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 19 Jul 2022

Authorised by: DR G FLOOD

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Scientist

Date: 20TH JULY 2022