

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**  
Revision: **2**  
Date: **09.01.14**  
Page: **1 of 1**

Prepared by: **Geoff Flood**  
Approved by: **Janet Bradbury**

Tracer iodination date: <sup>11</sup>12 August 2022

Kit lot number: 3INK321

Kit expiry date: 07 October 2022

125I - Insulin	Lot No. IT321R	Expiry Date: 07 OCT 2022
Assay Buffer	Lot No. MIAB136	Expiry Date: 19 OCT 2023
Anti-human IgG	Lot No. 179PG	Expiry Date: 05 NOV 2023
Calibrators	Lot No. IMS36E	Expiry Date: 15 JUN 2023
Negative Control	Lot No. IN130	Expiry Date: 10 OCT 2023
Kit Controls	Lot No. IMCa36 <sup>E</sup> / <del>IMCb36E</del>	Expiry Date: 15 JUN 2023

Total cpm in 25 µl 125I - Insulin 29335

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	1.0		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	3.5		
10 u/mL	15.6		
50 u/mL	56.1		
Control I	3.6	1.1	0.5 - 1.4
Control II	19.1	12.1	9 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 16 August 2022

Performed by: J Bradbury

Signature: J Bradbury

Position: Head of Operations

Authorised by: DR G Flood

Signature: G Flood

Position: Head of Quality Control

Date: 17<sup>th</sup> AUG 2022