

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood  
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b  
Revision: 2  
Date: 09.01.14  
Page 1 of 1

Tracer iodination date: 08 September 2022

Kit lot number: 3INK322

Kit expiry date: 04 November 2022

125I - Insulin	Lot No.	IT322R	Expiry Date:	04 NOV 2022
Assay Buffer	Lot No.	MIAB137	Expiry Date:	04 DEC 2023
Anti-human IgG	Lot No.	179PG	Expiry Date:	05 NOV 2023
Calibrators	Lot No.	IMS37	Expiry Date:	25 MAY 2024
Negative Control	Lot No.	IN130	Expiry Date:	10 OCT 2023
Kit Controls	Lot No.	IMCa37/IMCb37	Expiry Date:	25 MAY 2024


Total cpm in 25 µl 125I - Insulin 29332

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.1		
1 u/mL	3.69		
10 u/mL	15.6		
50 u/mL	52.3		
Control I	4.0	1.07	0.5 - 1.6
Control II	17.3	11.1	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.


Assay Date: 13 Sep 2022

Performed by: J DAVIES

Signature: 

Position: Scientist

Authorised by: DR G FLOOD

Signature: 

Position: Head of Quality Control

Date: 15<sup>TH</sup> SEP 2022