

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**
 Revision: **2**
 Date: **09.01.14**
 Page **1 of 1**

Prepared by: **Geoff Flood**
 Approved by: **Janet Bradbury**

Tracer iodination date: 06 October 2022

Kit lot number: 3INK323

Kit expiry date: 02 December 2022

125I - Insulin	Lot No.	IT323R	Expiry Date:	02 DEC 2022
Assay Buffer	Lot No.	MIAB137	Expiry Date:	04 DEC 2023
Anti-human IgG	Lot No.	180PG	Expiry Date:	27 DEC 2023
Calibrators	Lot No.	IMS37A	Expiry Date:	25 MAY 2024
Negative Control	Lot No.	IN130	Expiry Date:	10 OCT 2023
Kit Controls	Lot No.	IMCa37A/IMCb37A	Expiry Date:	25 MAY 2024

Total cpm in 25 μ l ¹²⁵I - Insulin 30040

<u>Sample</u>	<u>% ¹²⁵I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.3		
1 u/mL	4.3		
10 u/mL	16.5		
50 u/mL	51.3		
Control I	3.5	0.69	0.5 - 1.6
Control II	19.1	11.8	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 12 Oct 2022

Performed by: J DAVIES

Signature: 

Position: Scientist

Authorised by: DR G FLOOD

Signature: 

Position: Head of Quality Control

Date: 12th OCT 2022