

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**  
 Revision: **2**  
 Date: **09.01.14**  
 Page **1 of 1**

Prepared by: **Geoff Flood**  
 Approved by: **Janet Bradbury**

Tracer iodination date: 23 February 2023

Kit lot number: 3INK328


Kit expiry date: 21 April 2023


125I - Insulin	Lot No. <b>IT328R</b>	Expiry Date: <b>21 APR 2023</b>
Assay Buffer	Lot No. <b>MIAB140</b>	Expiry Date: <b>04 MAY 2024</b>
Anti-human IgG	Lot No. <b>181PG</b>	Expiry Date: <b>20 MAR 2024</b>
Calibrators	Lot No. <b>IMS37B</b>	Expiry Date: <b>25 MAY 2024</b>
Negative Control	Lot No. <b>IN130</b>	Expiry Date: <b>10 OCT 2023</b>
Kit Controls	Lot No. <b>IMCa37B/IMCb37B</b>	Expiry Date: <b>25 MAY 2024</b>

Total cpm in 25 µl 125I - Insulin 29188

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>0.9</u>		
INS-Ab calibrator:			
0.4 u/mL	2.1		
1 u/mL	4.0		
10 u/mL	14.8		
50 u/mL	51.7		
Control I	3.6	0.79	0.5 – 1.6
Control II	18.0	12.2	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 28 Feb 2023  
 Performed by: J DAVIES  
 Signature:   
 Position: Senior Scientist

Authorised by: DR G FLOOD  
 Signature:   
 Position: Head of Quality Control  
 Date: 1st MAR 2023