

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b

Revision: 2

Prepared by: Geoff Flood

Date: 09.01.14

Approved by: Janet Bradbury

Page 1 of 1

Tracer iodination date: 23 March 2023

Kit lot number: 3INK329

Kit expiry date: 19 May 2023

125I - Insulin	Lot No.	IT329R	Expiry Date:	19 MAY 2023
Assay Buffer	Lot No.	MIAB141	Expiry Date:	26 JUN 2024
Anti-human IgG	Lot No.	182PG	Expiry Date:	16 JUL 2024
Calibrators	Lot No.	IMS37B	Expiry Date:	25 MAY 2024
Negative Control	Lot No.	IN131	Expiry Date:	06 JUL 2024
Kit Controls	Lot No.	IMCa37B/IMCb37B	Expiry Date:	25 MAY 2024

Total cpm in 25 µl 125I - Insulin 29771

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>1.0</u>		
INS-Ab calibrator:			
0.4 u/mL	2.6		
1 u/mL	4.2		
10 u/mL	16.3		
50 u/mL	54.0		
Control I	4.1	0.93	0.5 - 1.6
Control II	21.0	13.1	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 28 Mar 2023

Authorised by: DR G Flood

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Senior Scientist

Date: 28th MAR 2023