

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**
 Revision: **2**
 Date: **09.01.14**
 Page **1 of 1**

Prepared by: **Geoff Flood**
 Approved by: **Janet Bradbury**

Tracer iodination date: 19 April 2023

Kit lot number: 3INK330

Kit expiry date: 15 June 2023

125I - Insulin	Lot No. IT330R	Expiry Date: 15 JUN 2023
Assay Buffer	Lot No. MIAB141	Expiry Date: 26 JUN 2024
Anti-human IgG	Lot No. 182PG	Expiry Date: 16 JUL 2024
Calibrators	Lot No. IMS37B	Expiry Date: 25 MAY 2024
Negative Control	Lot No. IN131	Expiry Date: 06 JUL 2024
Kit Controls	Lot No. IMCa37B/IMCb37B	Expiry Date: 25 MAY 2024

Total cpm in 25 µl 125I - Insulin 26800

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>0.9</u>		
INS-Ab calibrator:			
0.4 u/mL	2.5		
1 u/mL	3.9		
10 u/mL	16.8		
50 u/mL	53.2		
Control I	4.1	1.17	0.5 – 1.6
Control II	20.2	12.3	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.


Assay Date: 26 Apr 2023

Performed by: J DAVIES

Signature: 

Position: Senior Scientist

Authorised by: 

Signature: 

Position: Head of Quality Control

Date: 26th APR 2023