

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 15 June 2023

Kit lot number: 3INK332

Kit expiry date: 11 August 2023

125I - Insulin	Lot No.	IT332R	Expiry Date:	11 AUG 2023
Assay Buffer	Lot No.	MIAB143	Expiry Date:	23 AUG 2024
Anti-human IgG	Lot No.	183PG	Expiry Date:	18 OCT 2024
Calibrators	Lot No.	IMS37E	Expiry Date:	25 MAY 2024
Negative Control	Lot No.	IN131	Expiry Date:	06 JUL 2024
Kit Controls	Lot No.	IMCa37E/IMCb37E	Expiry Date:	25 MAY 2024

Total cpm in 25 µl 125I - Insulin 30376

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	1.0		
INS-Ab calibrator:			
0.4 u/mL	2.4		
1 u/mL	4.5		
10 u/mL	17.3		
50 u/mL	56.8		
Control I	5.0	1.4	0.5 - 1.6
Control II	19.6	11.4	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 20 Jun 2023

Performed by: A Huckle

Signature: A. Huckle

Position: Snr Technician

Authorised by: DR G Flood

Signature: [Signature]

Position: Head of Quality Control

Date: 26TH JUN 2023