

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b  
Revision: 2  
Date: 09.01.14  
Page 1 of 1

Prepared by: Geoff Flood  
Approved by: Janet Bradbury

Tracer iodination date: 07 September 2023

Kit lot number: 3INK335B

Kit expiry date: 03 November 2023

125I - Insulin	Lot No.	IT335R	Expiry Date:	03 NOV 2023
Assay Buffer	Lot No.	MIAB146	Expiry Date:	01 JAN 2025
Anti-human IgG	Lot No.	184PG	Expiry Date:	26 DEC 2024
Calibrators	Lot No.	IMS37E	Expiry Date:	25 MAY 2024
Negative Control	Lot No.	IN131	Expiry Date:	06 JUL 2024
Kit Controls	Lot No.	IMCa37E/IMCb37E	Expiry Date:	25 MAY 2024

Total cpm in 25 µl 125I - Insulin 30143

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.1		
1 u/mL	3.5		
10 u/mL	15.9		
50 u/mL	53.0		
Control I	3.9	1.4	0.5 - 1.6
Control II	19.7	12.5	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 12 Sep 2023

Authorised by: DR G FLOOD

Performed by: J DAVIES

Signature: [Signature]

Signature: [Signature]

Position: Head of Quality Control

Position: Senior Scientist

Date: 12<sup>TH</sup> SEP 2023