

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**  
Revision: **2**  
Date: **09.01.14**  
Page **1 of 1**

Prepared by: **Geoff Flood**  
Approved by: **Janet Bradbury**

Tracer iodination date: 22 February 2024

Kit lot number: 3INK341


Kit expiry date: 19 April 2024


125I - Insulin	Lot No.	IT341R	Expiry Date:	19 APR 2024
Assay Buffer	Lot No.	MIAB149	Expiry Date:	05 MAY 2025
Anti-human IgG	Lot No.	188PG	Expiry Date:	14 JUL 2025
Calibrators	Lot No.	IMS38A	Expiry Date:	26 FEB 2025
Negative Control	Lot No.	IN131	Expiry Date:	06 JUL 2024
Kit Controls	Lot No.	IMCa38A/IMCb38A	Expiry Date:	26 FEB 2025

Total cpm in 25 µl 125I - Insulin 26620

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	<u>0.7</u>		
INS-Ab calibrator:			
0.4 u/mL	2.3		
1 u/mL	4.4		
10 u/mL	18.9		
50 u/mL	60.6		
Control I	4.2	0.91	0.5 – 1.7
Control II	22.4	12.0	8 – 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 28 Feb 2024  
Performed by: J DAVIES  
Signature:   
Position: Senior Scientist

Authorised by: DR G Flood  
Signature:   
Position: Head of Quality Control  
Date: 29<sup>th</sup> FEB 2024