

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b  
Revision: 2  
Date: 09.01.14  
Page 1 of 1

Prepared by: Geoff Flood  
Approved by: Janet Bradbury

Tracer iodination date: 20 February 2025

Kit lot number: 3INK354B


Kit expiry date: 18 April 2025


125I - Insulin	Lot No.	IT354R	Expiry Date:	18 APRIL 2025
Assay Buffer	Lot No.	MIAB159	Expiry Date:	06 APR 2026
Anti-human IgG	Lot No.	196PG	Expiry Date:	03 AUG 2026
Calibrators	Lot No.	IMS39B	Expiry Date:	27 FEB 2026
Negative Control	Lot No.	IN132	Expiry Date:	04 JUL 2025
Kit Controls	Lot No.	IMCa39B/IMCb39B	Expiry Date:	27 FEB 2026

Total cpm in 25 µl 125I - Insulin 27671

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.8		
INS-Ab calibrator:			
0.4 u/mL	2.6		
1 u/mL	4.1		
10 u/mL	20.1		
50 u/mL	62.0		
Control I	4.1	1.0	0.5 - 1.5
Control II	25.2	12.9	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 26 Feb 2025  
Performed by: J DAVIES  
Signature:   
Position: Senior Scientist

Authorised by: DR G Flood  
Signature:   
Position: Head of Quality Control  
Date: 26<sup>th</sup> FEB 2025