

Title: **Quality Control Record Sheet  
Insulin Ab 3 QC Data**

Document: **wi/qcrs/0.08b**

Revision: **2**

Prepared by: **Geoff Flood**

Date: **09.01.14**

Approved by: **Janet Bradbury**

Page **1 of 1**

Tracer iodination date: 20 March 2025

Kit lot number: 3INK355

Kit expiry date: 16 May 2025

125I - Insulin	Lot No. IT355R	Expiry Date: 16 MAY 2025
Assay Buffer	Lot No. MIAB161	Expiry Date: 21 JUN 2026
Anti-human IgG	Lot No. 196PG	Expiry Date: 03 AUG 2026
Calibrators	Lot No. IMS39E	Expiry Date: 27 FEB 2026
Negative Control	Lot No. IN132	Expiry Date: 04 JUL 2025
Kit Controls	Lot No. IMCa39E/IMCb39E	Expiry Date: 27 FEB 2026

Total cpm in 25 µl 125I - Insulin 30735

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.6		
INS-Ab calibrator:			
0.4 u/mL	2.2		
1 u/mL	3.6		
10 u/mL	16.1		
50 u/mL	58.2		
Control I	3.2	0.73	0.5 - 1.5
Control II	20.7	12.7	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 25 Mar 2025

Authorised by: SARAH NUTE

Performed by: J DAVIES

Signature: SRNute

Signature: 

Position: pp Head of Quality Control

Position: Senior Scientist

Date: 25 MAR 2025