

Title: **Quality Control Record Sheet
Insulin Ab 3 QC Data**

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 02 October 2025

Kit lot number: 3INK362


Kit expiry date: 28 November 2025


125I - Insulin	Lot No.	IT362R	Expiry Date:	28 NOV 2025
Assay Buffer	Lot No.	MIAB168	Expiry Date:	28 JAN 2027
Anti-human IgG	Lot No.	199PG	Expiry Date:	03 FEB 2027
Calibrators	Lot No.	IMS40B	Expiry Date:	26 FEB 2027
Negative Control	Lot No.	IN133	Expiry Date:	16 JUL 2026
Kit Controls	Lot No.	IMCa40B/IMCb40B	Expiry Date:	26 FEB 2027

Total cpm in 25 μ l 125I - Insulin 27805

<u>Sample</u>	<u>% 125I - Insulin bound</u>	<u>u/mL</u>	<u>Range u/mL</u>
Negative control	0.7		
INS-Ab calibrator:			
0.4 u/mL	2.5		
1 u/mL	4.1		
10 u/mL	19.1		
50 u/mL	58.1		
Control I	4.2	1.06	0.45 - 1.45
Control II	22.8	12.3	6 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 07 Oct 2025
Performed by: J DAVIES
Signature: 
Position: Senior Scientist

Authorised by: DR G Flood
Signature: 
Position: Head of Quality Control
Date: 7th Oct 2025