

Title: **Quality Control Record Sheet**  
**Insulin Ab 3 QC Data**

Document: wi/qcrs/0.08b

Revision: 2

Date: 09.01.14

Page 1 of 1

Prepared by: Geoff Flood

Approved by: Janet Bradbury

Tracer iodination date: 27 November 2025

Kit lot number: 3INK364

Kit expiry date: 23 January 2026

125I - Insulin	Lot No. IT364R	Expiry Date: 23 JAN 2026
Assay Buffer	Lot No. MIAB169	Expiry Date: 09 MAR 2027
Anti-human IgG	Lot No. 200PG	Expiry Date: 20 FEB 2027
Calibrators	Lot No. IMS40D	Expiry Date: 26 FEB 2027
Negative Control	Lot No. IN133	Expiry Date: 16 JUL 2026
Kit Controls	Lot No. IMCa40D/IMCb40D	Expiry Date: 26 FEB 2027

Total cpm in 25 µl 125I - Insulin 28251

<u>Sample</u>	<u>% 125I - Insulin</u> <u>bound</u>	<u>u/mL</u>	<u>Range</u> <u>u/mL</u>
Negative control	0.8		
INS-Ab calibrator:			
0.4 u/mL	2.6		
1 u/mL	4.1		
10 u/mL	17.9		
50 u/mL	55.4		
Control I	4.1	1.00	0.45 - 1.45
Control II	20.9	11.9	6 - 18

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 02 Dec 2025

Performed by: J DAVIES

Signature: 

Position: Senior Scientist

Authorised by: SARAH NUTE

Signature: 

Position: pp Head of Quality Control

Date: 03 DEC 2025