Quality Control Record Sheet ¹²⁵I-labelled MuSK QC Data Title:

G FLOOD Prepared by:

Approved by: J BRADBURY Document: Revision:

wi/qcrs/0.30a

Issue date:

20.10.2021

Page

1 of 1

Kit Lot Number:	KMK242		Kit Expiry Date: 13 NOVEMBER 2023	
¹²⁵ I-labelled MuSK:	MK242		Expiry Date: <u>13 NOVEMBER 2023</u>	
Reconstitution Buffer:	170B		Expiry Date: 25 APRIL 2024	
Anti Human IgG:	181Gd		Expiry Date: 24 JUNE 2024	
Precipitation Enhancer:	PE190		Expiry Date: 15 MAY 2024	
Wash Solution:	431W		Expiry Date: 02 JANUARY 2024	
Negative Control:	MKN41f		Expiry Date: 24 MARCH 2024	
Positive Control:	MKP41n		Expiry Date: 24 MARCH 2024	
Specific activity of 125I-labelled MuSK:		1692	Ci/mmol (30.8 μCi/μg)	
Total cpm in 50 μL of labelled MuSK:		44842		
MuSK labelling date: <u>18 SEPTEMBER 2023</u>				
% Counter Efficiency:	72.3			

125 Decay Table:

QC date	Decay factor		
(weeks after labelling date)			
up to 1 week	1.0		
+1 - 2 weeks	1.1		
+2 - 3 weeks	1.2		
+3 - 4 weeks	1.3		
+4 - 5 weeks	1.4		
+ 5 - 6 weeks	1.5		
+ 6 - 7 weeks	1.6		
+ 7 - 8 weeks	1.75		

Positive Control	Mean cpm	Nmol/L	Range (nmol/L)
Negative Control	301	-	-
Positive Control	9762	0.6862	0.77 ± 0.38

Materials of human origin used in the manufacture of this product have been tested and found nonreactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 21 SEPTEMBER 2023 Performed by: _____ Jonathon Davies

Signature: _____

Position: Senior Scientist

Authorised by:

Signature: _

Position: Head of Quality Control

Date: 21 ST

SEP