Title:

Quality Control Record Sheet <sup>125</sup>I-labelled MuSK QC Data

Prepared by:

G FLOOD

Approved by: J BRADBURY

Document:

wi/qcrs/0.30a

Revision: Issue date:

20.10.2021

Page

1 of 1

Kit Lot Number:	KMK218	Kit Expiry Date: 02 MAY 2022		
<sup>125</sup> I-labelled MuSK:	MK218	Expiry Date: 02 MAY 2022		
Reconstitution Buffer:	165B	Expiry Date: 01 SEPTEMBER 2022		
Anti Human IgG:	179Gw	Expiry Date: 20 NOVEMBER 2022		
Precipitation Enhancer:	PE181	Expiry Date: 27 OCTOBER 2022		
Wash Solution:	406W	Expiry Date: 12 MAY 2022		
Negative Control:	MKN39a	Expiry Date: 13 SEPTEMBER 2022		
Positive Control:	MKP38f	Expiry Date: 13 SEPTEMBER 2022		
Specific activity of <sup>125</sup> I-labelled MuSK:1878		Ci/mmol ( 34.1 μCi/μg)		
Total cpm in 50 μL of labelled MuS				
MuSK labelling date: 07 March 20				
% Counter Efficiency: 72.3				

125 Decay Table:

QC date	Decay factor		
(weeks after labelling date)			
up to 1 week	1.0		
+1 - 2 weeks	1.1		
+2 - 3 weeks	1.2		
+3 - 4 weeks	1.3		
+4 - 5 weeks	1.4		
+ 5 - 6 weeks	1.5		
+ 6 - 7 weeks	1.6		
+ 7 - 8 weeks	1.75		

Positive Control	Mean cpm	Nmol/L	Range (nmol/L)
Negative Control	418	-	-
Positive Controll	11409	0.7293	0.66 ± 0.33

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date:	10 March 2022	
Performed by: _	Jonathon Davies	
Signature:		

Position: Head of 0

Authorised by: DR G

Head of Quality Control

Position: Scientist

Date: 11TH MAR 2022