

Title: **Quality Control Record Sheet**
MuSK Ab Assay QC Data

Prepared by: G FLOOD
Approved by: J BRADBURY

Document: wi/qcrs/0.30a
Revision: 0
Issue date: 20.10.2021
Page 1 of 1

Kit Lot Number: KMK273
¹²⁵I-labelled MuSK: MK273
Reconstitution Buffer: 177B
Anti Human IgG: 185Gt
Precipitation Enhancer: PE201
Wash Solution: 469W
Negative Control: MKN44k
Positive Control: MKP44t

Kit Expiry Date: 05 JANUARY 2026
Expiry Date: 05 JANUARY 2026
Expiry Date: 14 MARCH 2026
Expiry Date: 16 AUGUST 2026
Expiry Date: 05 JUNE 2026
Expiry Date: 04 JUNE 2026
Expiry Date: 16 MAY 2026
Expiry Date: 16 MAY 2026

Specific activity of ¹²⁵I-labelled MuSK: 1921 Ci/mmol (34.9 μ Ci/ μ g)

Total cpm in 50 μ L of labelled MuSK: 51520

MuSK labelling date: 10 NOVEMBER 2025

% Counter Efficiency: 73.4

¹²⁵I Decay Table:

QC date (weeks after labelling date)	Decay factor
up to 1 week	1.0
+1 - 2 weeks	1.1
+2 - 3 weeks	1.2
+3 - 4 weeks	1.3
+4 - 5 weeks	1.4
+ 5 - 6 weeks	1.5
+ 6 - 7 weeks	1.6
+ 7 - 8 weeks	1.75

Positive Control	Mean cpm	Nmol/L	Range (nmol/L)
Negative Control	349	-	-
Positive Control	14724	0.9325	0.77 $\pm$ 0.38

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 13 NOVEMBER 2025
Performed by: Jonathon Davies
Signature: [Signature]
Position: Senior Scientist

Authorised by: DR G Flood
Signature: [Signature]
Position: Head of Quality Control
Date: 13TH NOV 2025