

Title: **Quality Control Record Sheet
MuSK Ab Assay QC Data**

Prepared by: G FLOOD
Approved by: J BRADBURY

Document: wi/qcrs/0.30a
Revision: 0
Issue date: 20.10.2021
Page 1 of 1

Kit Lot Number: KMK278A
¹²⁵I-labelled MuSK: MK278
Reconstitution Buffer: 178B
Anti Human IgG: 185Gy
Precipitation Enhancer: PE203
Wash Solution: 477W
Negative Control: MKN45a
Positive Control: MKP45e

Kit Expiry Date: 25 MAY 2026
Expiry Date: 25 MAY 2026
Expiry Date: 31 JULY 2026
Expiry Date: 05 JANUARY 2027
Expiry Date: 23 OCTOBER 2026
Expiry Date: 17 DECEMBER 2026
Expiry Date: 05 OCTOBER 2026
Expiry Date: 05 OCTOBER 2026

Specific activity of ¹²⁵I-labelled MuSK: 1881 Ci/mmol (34.2 μ Ci/ μ g)

Total cpm in 50 μ L of labelled MuSK: 51448

MuSK labelling date: 30 MARCH 2026

% Counter Efficiency: 75.8

¹²⁵I Decay Table:

QC date (weeks after labelling date)	Decay factor
up to 1 week	1.0
+1 - 2 weeks	1.1
+2 - 3 weeks	1.2
+3 - 4 weeks	1.3
+4 - 5 weeks	1.4
+ 5 - 6 weeks	1.5
+ 6 - 7 weeks	1.6
+ 7 - 8 weeks	1.75

Positive Control	Mean cpm	Nmol/L	Range (nmol/L)
Negative Control	439	-	-
Positive Control	16391	1.0079	0.82 \pm 0.41

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 02 APRIL 2026
Performed by: Jonathon Davies
Signature: [Signature]
Position: Senior Scientist

Authorised by: DR G Flood
Signature: [Signature]
Position: Head of Quality Control
Date: 2nd APR 2026