Title:

Quality Control Record Sheet RBA AChRAb Assay QC data

Prepared by: Geoff Flood Approved by: Rachael Price Document: Revision:

wi/qcrs/0.14a

Date:

6 Sep 2018

Page

2 of 2

Kit lot no:

KRBA347

Expiry date:

13 May 2024

Assay date (weeks after receptor labelling)	Decay factor (A)
Up to 1 week	1.0
+ 1 – 2 weeks	1.1
+ 2 – 3 weeks	1.2
+ 3 – 4 weeks	1.3
+ 4 – 5 weeks	1.4
+ 5 – 6 weeks	1.5
+ 6 – 7 weeks	1.6
+ 7 – 8 weeks	1.75
+ 8 – 9 weeks	1.9

Materials of human origin used in the manufacture of this product have been tested and found nonreactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay date: Performed by: 14 Mar 2024

Signature: Position:

A Lee

Principal Technician

Authorised by:

Signature:

Position:

Date:

Head of Quality Control