

Title: **Quality Control Record Sheet**
RBA AChRAb Assay QC data

Document: **wi/qcrs/0.14a**

Revision: **7**

Prepared by: Geoff Flood

Date: **6 Sep 2018**

Approved by: Rachael Price

Page **2 of 2**

Kit lot no:

KRBA348

Expiry date:


10 Jun 2024

Assay date (weeks after receptor labelling)	Decay factor (A)
Up to 1 week	1.0
+ 1 – 2 weeks	1.1
+ 2 – 3 weeks	1.2
+ 3 – 4 weeks	1.3
+ 4 – 5 weeks	1.4
+ 5 – 6 weeks	1.5
+ 6 – 7 weeks	1.6
+ 7 – 8 weeks	1.75
+ 8 – 9 weeks	1.9

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay date: **11 Apr 2024**

Performed by: **A Lee**

Signature: 

Position: **Principal Technician**

Authorised by: **DR G FLOOD**

Signature: 

Position: **Head of Quality Control**

Date: **12TH APR 2024**