Title:

Quality Control Record Sheet RBA AChRAb Assay QC data

Prepared by: Geoff Flood Approved by: Rachael Price

Date: Page

Document:

Revision:

wi/qcrs/0.14a

6 Sep 2018 2 of 2

Kit lot no:

KRBA358

Expiry date:

17 Mar 2025

Assay date (weeks after receptor labelling)	Decay factor (A)
Up to 1 week	1.0
+ 1 – 2 weeks	1.1
+ 2 – 3 weeks	1.2
+ 3 – 4 weeks	1.3
+ 4 – 5 weeks	1.4
+ 5 – 6 weeks	1.5
+ 6 – 7 weeks	1.6
+ 7 – 8 weeks	1.75
+ 8 – 9 weeks	1.9

Materials of human origin used in the manufacture of this product have been tested and found nonreactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay date:

16 Jan 2025

Authorised by:

Performed by:

A Lee

Signature:

Signature:

Position:

Head of Quality Control

Position:

Principal Technician

Date: