

Title: **Quality Control Record Sheet
AChRab Standard Curve QC Data**

Document: **wi/qcrs/0.14SCb**

Revision: **3**

Prepared by: Geoff Flood

Date: **6 Sep 2018**

Approved by: Rachael Price

Page **1 of 1**

Re: AChRab Standard Curve (lot no. ASC20c; expiry date. 02 Jul 2023)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA328** are as follows:


Standard	Actual value at QC
1	0.32 nmol/litre
2	1.4 nmol/litre
3	4.3 nmol/litre
4	9.2 nmol/litre
Value of positive control 116PCr read off standard curve = 5.0 nmol/litre	
(Range: 2.8-6.4 nmol/litre)	


Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 29 Sep 2022

Authorised by: SARAH NUTE

Performed by: A Lee

Signature: 

Signature: 

Position: pp Head of Quality Control

Position: Principal Technician

Date: 30 SEP 2022