

Title: **Quality Control Record Sheet
AChRAb Standard Curve QC Data**

Prepared by: Geoff Flood
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**
Revision: **3**
Date: **6 Sep 2018**
Page **1 of 1**

Re: AChRAb Standard Curve (lot no. ASC20h; expiry date. 19 Nov 2023)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA333** are as follows:

Standard	Actual value at QC
1	0.31 nmol/litre
2	1.0 nmol/litre
3	3.8 nmol/litre
4	8.7 nmol/litre
Value of positive control 116PCx read off standard curve = 4.7 nmol/litre (Range: 2.8-6.4 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 16 Feb 2023

Performed by: A Lee

Signature: 

Position: Principal Technician

Authorised by: SARAH NUTE

Signature: 

Position: pp Head of Quality Control

Date: 17 FEB 2023