

Title: **Quality Control Record Sheet**  
**AChRab Standard Curve QC Data**

Prepared by: Geoff Flood  
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**  
Revision: **3**  
Date: **6 Sep 2018**  
Page **1 of 1**

**Re: AChRab Standard Curve (lot no. ASC20p; expiry date. 11 Mar 2024 )**

*To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.*

*No <sup>125</sup>I decay correction is needed.*

Our QC data for the standard curve, obtained with AChR lot **RBA337** are as follows:

Standard	Actual value at QC
1	0.31 nmol/litre
2	1.0 nmol/litre
3	3.4 nmol/litre
4	8.3 nmol/litre
Value of positive control <b>116PCac</b> read off standard curve = 4.2 nmol/litre (Range: 2.8-6.4 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date: 08 Jun 2023

Performed by: A Lee

Signature: 

Position: Principal Technician

Authorised by: DR G FLOOD

Signature: 

Position: Head of Quality Control

Date: 7<sup>TH</sup> JUN 2023