

Title: **Quality Control Record Sheet
AChRAb Standard Curve QC Data**

Document: **wi/qcrs/0.14SCb**
 Revision: **3**
 Date: **6 Sep 2018**
 Page: **1 of 1**

Prepared by: Geoff Flood
 Approved by: Rachael Price

Re: AChRAb Standard Curve (lot no. ASC20u; expiry date. 03 Jun 2024)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA340** are as follows:

Standard	Actual value at QC
1	0.29 nmol/litre
2	1.1 nmol/litre
3	4.6 nmol/litre
4	8.8 nmol/litre
Value of positive control 116PCaf read off standard curve = 4.7 nmol/litre	
(Range: 2.8-6.4 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 01 Sep 2023

Performed by: A.Lee

Signature: *pp. A.Lee*

Position: Principal Technician

Authorised by: *DR G flood*

Signature: *[Signature]*

Position: Head of Quality Control

Date: 4th SEP 2023