

Title: **Quality Control Record Sheet**  
**AChRab Standard Curve QC Data**

Prepared by: Geoff Flood  
 Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**  
 Revision: **3**  
 Date: **6 Sep 2018**  
 Page: **1 of 1**

**Re: AChRab Standard Curve (lot no. ASC20x; expiry date. 26 Aug 2024)**


*To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.*


*No <sup>125</sup>I decay correction is needed.*

Our QC data for the standard curve, obtained with AChR lot **RBA343** are as follows:

Standard	Actual value at QC
1	0.26 nmol/litre
2	1.2 nmol/litre
3	4.0 nmol/litre
4	8.5 nmol/litre
Value of positive control <b>116PCak</b> read off standard curve = 4.7 nmol/litre	
(Range: 2.8-6.4 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 23 Nov 2023  
 Performed by: A Lee  
 Signature:   
 Position: Principal Technician

Authorised by: DR G Flood  
 Signature:   
 Position: Head of Quality Control  
 Date: 24<sup>th</sup> NOV 2023