

Title: **Quality Control Record Sheet  
AChRAb Standard Curve QC Data**

Prepared by: Geoff Flood  
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**  
Revision: **3**  
Date: **6 Sep 2018**  
Page **1 of 1**

**Re: AChRAb Standard Curve (lot no. ASC20ab; expiry date. 18 Nov 2024)**


To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.



No <sup>125</sup>I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA346** are as follows:

Standard	Actual value at QC
1	0.23 nmol/litre
2	0.83 nmol/litre
3	3.1 nmol/litre
4	8.3 nmol/litre
Value of positive control <b>118Pb</b> read off standard curve = 3.8 nmol/litre	
(Range: 2.5-5.9 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 15 Feb 2024  
Performed by: R. Price  
Signature:   
Position: Principal Scientist

Authorised by:   
Signature:   
Position: Head of Quality Control  
Date: 15<sup>TH</sup> FEB 2024