

Title: **Quality Control Record Sheet  
AChRab Standard Curve QC Data**

Document: **wi/qcrs/0.14SCb**  
 Revision: **3**  
 Date: **6 Sep 2018**  
 Page **1 of 1**

Prepared by: Geoff Flood  
 Approved by: Rachael Price

**Re: AChRab Standard Curve (lot no. ASC20m; expiry date. 16 Jan 2024 )**


*To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.*



*No <sup>125</sup>I decay correction is needed.*

Our QC data for the standard curve, obtained with AChR lot **RBA335** are as follows:

Standard	Actual value at QC
1	0.28 nmol/litre
2	0.92 nmol/litre
3	3.5 nmol/litre
4	7.9 nmol/litre
Value of positive control <b>116PCaa</b> read off standard curve = 4.4 nmol/litre	
(Range: 2.8-6.4 nmol/litre)	

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 14 Apr 2023  
 Performed by: R.Price  
 Signature:   
 Position: Principal Scientist

Authorised by:   
 Signature:   
 Position: Head of Quality Control  
 Date: 17<sup>TH</sup> APR 2023