

Title: **Quality Control Record Sheet**
AChRAb Standard Curve QC Data

Prepared by: Geoff Flood
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**
Revision: **3**
Date: **6 Sep 2018**
Page **1 of 1**

Re: AChRAb Standard Curve (lot no. ASC21af; expiry date. 05 Apr 2027)


To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.


No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA377** are as follows:

Standard	Actual value at QC	
1	0.25	nmol/litre
2	0.90	nmol/litre
3	3.3	nmol/litre
4	8.5	nmol/litre
Value of positive control 118Pat read off standard curve = 4.3 nmol/litre		
(Range: 2.5-5.9 nmol/litre)		

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 02 Jul 2026
Performed by: A Lee
Signature: 
Position: Principal Technician

Authorised by: 
Signature: 
Position: Head of Quality Control
Date: 30 JULY 2026