

Title: **Quality Control Record Sheet**
AChRab Standard Curve QC Data

Prepared by: Geoff Flood
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**
Revision: **3**
Date: **6 Sep 2018**
Page **1 of 1**

Re: AChRab Standard Curve (lot no. ASC21w; expiry date. 23 Aug 2026)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA369** are as follows:

Standard	Actual value at QC	
1	0.30	nmol/litre
2	0.84	nmol/litre
3	3.4	nmol/litre
4	8.0	nmol/litre
Value of positive control 118Paf read off standard curve = 4.4 nmol/litre		
(Range: 2.5-5.9 nmol/litre)		

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 20 Nov 2025

Performed by: A Lee

Signature: 

Position: Principal Technician

Authorised by: DR G FLOOD

Signature: 

Position: Head of Quality Control

Date: 21st Nov 2025