

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: KTCT413

Kit expiry date: 26 Jun 2025

125I -TSH Tracer:	Lot No. TCT377	Expiry Date 26 Jun 2025
TSH Receptor Coated Tubes:	Lot No. 2RC1075	Expiry Date 09 Aug 2025
Calibrators:	Lot No. 2CTC288	Expiry Date 02 Dec 2026
Negative Control:	Lot No. 2CTC288N	Expiry Date 02 Dec 2026
Positive Control:	Lot No. 2CTC288P	Expiry Date 02 Dec 2026
Concentrated Wash Solution:	Lot No. TEW146	Expiry Date 14 Oct 2027
Start Buffer:	Lot No. 2aAD112	Expiry Date 02 Aug 2026

Total counts: 87791 cpm

Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	13.61	17	
2	2	11.56	29	
3	8	5.04	69	
4	40	1.33	92	
CONTROL	% bound	% inhibition	IU/L	Range
I (Negative)	16.31	-	-	
II (Positive)	8.53	48	3.9	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 24 Apr 2025

Performed by: Claire Fisher

Signature: pp W. Dowler

Position: Laboratory Technician

Authorised by: SARAH NUTE

Signature: SRNute

Position: pp Head of Quality Control

Date: 25 APR 2025