

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: KTCT419B

Kit expiry date: 10 Dec 2025

125I –TSH Tracer:	Lot No.	TCT383	Expiry Date	10 Dec 2025
TSH Receptor Coated Tubes:	Lot No.	2RC1090	Expiry Date	03 Feb 2026
Calibrators:	Lot No.	2CTC294	Expiry Date	26 Jun 2027
Negative Control:	Lot No.	2CTC294N	Expiry Date	26 Jun 2027
Positive Control:	Lot No.	2CTC294P	Expiry Date	26 Jun 2027
Concentrated Wash Solution:	Lot No.	TEW146	Expiry Date	14 Oct 2027
Start Buffer:	Lot No.	2aAD112	Expiry Date	02 Aug 2026

Total counts: 85662 cpm


Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
Calibrator				
1	1	12.77	15	
2	2	11.54	23	
3	8	6.05	60	
4	40	2.21	85	
CONTROL	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	15.00	-	-	
II (Positive)	9.54	36	3.5	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.


Assay Date: 08 Oct 2025

Performed by: Claire Fisher

Signature: 

Position: Laboratory Technician

Authorised by: SARAH NUTE

Signature: 

Position: pp Head of Quality Control

Date: 10 OCT 2025