

Title: **Quality Control Record Sheet**
TRAb Coated Tubes QC Data

Prepared by: G Flood
Approved by: W Dowler

Document: wi/qcrs/0.23
Revision: 5
Date: 11.05.21
Page 1 of 1

Kit Lot Number: KTCT420A

Kit expiry date: 07 Jan 2026

125I –TSH Tracer:	Lot No. TCT384	Expiry Date 07 Jan 2026
TSH Receptor Coated Tubes:	Lot No. 2RC1091	Expiry Date 25 Feb 2026
Calibrators:	Lot No. 2CTC295	Expiry Date 26 Jun 2027
Negative Control:	Lot No. 2CTC295N	Expiry Date 26 Jun 2027
Positive Control:	Lot No. 2CTC295P	Expiry Date 26 Jun 2027
Concentrated Wash Solution:	Lot No. TEW148	Expiry Date 11 Aug 2028
Start Buffer:	Lot No. 2aAD112	Expiry Date 02 Aug 2026

Total counts: 85710 cpm

Temp. 21 °C

	<u>IU/L</u>	<u>% bound</u>	<u>% inhibition</u>	
<u>Calibrator</u>				
1	1	12.45	14	
2	2	11.23	23	
3	8	6.49	55	
4	40	2.56	82	
<u>CONTROL</u>	<u>% bound</u>	<u>% inhibition</u>	<u>IU/L</u>	<u>Range</u>
I (Negative)	14.51	-	-	
II (Positive)	9.03	38	4.0	3.9 +/- 1.2 IU/L

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 05 Nov 2025

Performed by: Wendy Dowler

Signature: pp sarah nute

Position: Section Leader

Authorised by: SARAH NUTE

Signature: SRNute

Position: pp Head of Quality Control

Date: 06 NOV 2025